

Xenium In Situ for FFPE – Deparaffinization & Decrosslinking

Introduction

Xenium In Situ for FFPE is designed to measure mRNA in tissue sections derived from formalin fixed & paraffin embedded (FFPE) tissue samples and requires a Xenium slide with intact tissue sections as input. This protocol outlines deparaffinization and decrosslinking of FFPE tissues for use with 10x Genomics Xenium In Situ Gene Expression protocols. Deparaffinized and decrosslinked tissue sections are inputs for the downstream Xenium In Situ Gene Expression - Probe Hybridization, Ligation & Amplification workflow.

Additional Guidance

Consult the Xenium In Situ for FFPE - Tissue Preparation Guide (Document CG000578) for complete information on sectioning FFPE tissue blocks and placing sections on Xenium slides. After completing this Demonstrated Protocol (CG000580), proceed with the Xenium In Situ Gene Expression - Probe Hybridization, Ligation & Amplification User Guide (CG000582).













Contents

Xenium In Situ Gene Expression Reagent Kits	3
Xenium Slides & Sample Prep Reagents (2 slides, 2 rxns) PN-1000460	3
Xenium Decoding Consumables (1 run, 2 slides) PN-1000487	4
Xenium Instrument Accessory Kit Module A PN-1000530	4
Xenium Cassette Kit - (2 cassettes) PN-1000566	5
Recommended Thermal Cyclers	6
Workflow Overview	7
Specific Reagents & Consumables	8
Tips & Best Practices	10
Icons	10
Pipette Calibration	10
Xenium Slide	11
Xenium Slide Handling	12
Slide Storage	13
Xenium Cassette	14
Xenium Cassette Assembly	16
Xenium Cassette Removal	17
Reagent Addition to Wells	18
Reagent Removal from Wells	18
Xenium Cassette Lid Application & Removal	19
Slide Incubation Guidance	20
Processing a Single Xenium Slide	22
Tissue Detachment on Xenium Slides	22
Protocol Steps & Timing	23
1. Deparaffinization & Decrosslinking	24
1.0 Overview	24
Protocol Overview	25
Get Started - Deparaffinization & Decrosslinking	27
1.1 Preparation - Buffers	29
1.2 Deparaffinization	32
1.3 Cassette Assembly	34
1.4 Decrosslinking	35
Troubleshooting	37
Document Revision Summary	39

Xenium In Situ Gene Expression Reagent Kits





Refer to SDS for handling and disposal information


Xenium Slides & Sample Prep Reagents (2 slides, 2 rxns) PN-1000460

Xenium Slides & Sample Prep Reagents (2 slides, 2 rxns), PN-1000460 Store at -20°C		
	#	PN
 Xenium Probe Hybridization Buffer	1	2000390
 Xenium Post Hybridization Wash Buffer	1	2000395
 Xenium Ligation Buffer	1	2000391
 Xenium Ligation Enzyme A	1	2000397
 Xenium Ligation Enzyme B	1	2000398
 Xenium Amplification Mix	1	2000392
 Xenium Amplification Enzyme	1	2000399
 Reducing Agent B	1	2000087
 Xenium Autofluorescence Mix	1	2000753
 Xenium FFPE Tissue Enhancer	1	2000798
 Xenium Nuclei Staining Buffer	1	2000762
 Perm Enzyme B	1	3000553
Xenium Slides (2 pack)	1	3000941

Only Xenium FFPE Tissue Enhancer (PN-2000798) and Perm Enzyme B (PN-3000553) are needed for this workflow.

Xenium Decoding Consumables (1 run, 2 slides) PN-1000487


Xenium Decoding Consumables (1 run, 2 slides), PN-1000487 <i>Store at ambient temperature</i>		
	#	PN
Xenium Cassette Kit (2 cassettes + 16 lids)	1	1000566
Extraction Tip	1	2000757
Pipette Tips	1	3000866
Xenium Buffer Cap	1	3000949
Xenium Objective Wetting Consumable	1	2000749
 Deionized Water (bottle)	1	3001198
 Xenium Sample Wash Buffer A (bottle)	1	3001199
 Xenium Sample Wash Buffer B (bottle)	1	3001200
 Xenium Probe Removal Buffer (bottle)	1	3001201



Only the Xenium Cassette Kit (2 cassettes + 16 lids) (PN-1000566) is needed for this workflow.


Xenium Instrument Accessory Kit Module A PN-1000530

Xenium Instrument Accessory Kit Module A PN-1000530 <i>Store at ambient temperature</i>		
	#	PN
Waste Bottle	1	3000955
Xenium Waste Tip Tray	1	3000957
Xenium Thermocycler Adaptor	1	3000954



Only the Xenium Thermocycler Adaptor (PN-3000954) is needed for this workflow.

Xenium Cassette Kit - (2 cassettes) PN-1000566

Xenium Cassette Kit (2 cassettes) PN-1000566 <i>Store at ambient temperature</i>		
	#	PN
Xenium Cassette lids (16 ct)	1	3001046
Xenium Cassettes (2 pack)	1	3000951
		

Purchase the Xenium Cassette Kit (2 cassettes) (PN-1000566) for additional cassettes as needed.

Recommended Thermal Cyclers

Supplier	Description	Part Number
Bio-Rad	C1000 Touch Thermal Cycler with 96-Deep Well Reaction Module	1851197
Thermo Fisher Scientific	Veriti 96-Well Thermal Cycler (discontinued)	4375786
Analytik Jena	Biometra TAdvanced 96 SG with 96-well block (silver, 0.2 mL) and gradient function	846-x-070-241 (where x=2 for 230 V; 4 for 115 V; 5 for 100 V, 50-60 Hz)
VWR	Gradient thermal cycler, XT ⁹⁶ Gradient, with 96-well gradient block and standard lid	76452-153
Marshall Scientific	MJ Research PTC-200 Thermal Cycler	05434-05

Workflow Overview

Tissue Preparation Guide

Section FFPE tissue onto slides
Demonstrated Protocol CG000578



Deparaffinization + Decrosslinking

Deparaffinize and decrosslink tissue
Demonstrated Protocol CG000580



Probe Hybridization, Ligation & Amplification

For FF & FFPE Samples on Xenium Slides
User Guide CG000582



Decoding & Imaging

Load Xenium Cassette onto Xenium Analyzer
User Guide CG000584



Optional:

H&E Staining

H&E Stain post-Xenium run
Demonstrated Protocol CG000613

Visit the [10x Genomics Support website](https://www.10xgenomics.com/support) for the most current documentation.

Specific Reagents & Consumables

FFPE Tissue Sections: Deparaffinization & Decrosslinking

The listed items have been tested by 10x Genomics and perform optimally with the assay.

Substituting materials may adversely affect system performance. For items with multiple options listed, choose option based on availability and preference. Refer to the manufacturer's website for regional part numbers.

For FFPE Tissue Sections: Deparaffinization & Decrosslinking				
Item		Description	Vendor	Part Number
<input type="checkbox"/>	Xylene	Xylene, Reagent Grade	Millipore Sigma	214736
	or	Xylene, Histological Grade	Millipore Sigma	534056
	Neo-clear	Neo-clear Xylene Alternative Substitute	Millipore Sigma	1098435000
<input type="checkbox"/>	Ethanol	Ethyl Alcohol, 200 Proof, anhydrous	Millipore Sigma	E7023
		Ethanol absolute ≥99.5%, TechniSolv, pure (Europe)	VWR	83813.360DP
<input type="checkbox"/>	Nuclease-free Water	Nuclease-free Water (not DEPC-treated)	Thermo Fisher Scientific	AM9932/ AM9937
<input type="checkbox"/>	PBS	PBS - Phosphate Buffered Saline (10X) pH 7.4, RNase-free	Thermo Fisher Scientific	AM9624
<input type="checkbox"/>	Urea	Urea Solution, 8M	Millipore Sigma	51457
<input type="checkbox"/>	10% Tween-20	Tween 20 Surfact-Amps Detergent Solution (10% solution) (<i>not 100% Tween diluted to 10%</i>)	Thermo Fisher Scientific	28320
<input type="checkbox"/>	Perm Enzyme B	Perm Enzyme B	10x Genomics	3000553
<input type="checkbox"/>	Xenium FFPE Tissue Enhancer	Xenium FFPE Tissue Enhancer	10x Genomics	2000798
<input type="checkbox"/>	Forceps	Tweezers, 4" Wafer Handling	Excelta Corp	491P-SA-PI
<input type="checkbox"/>	Staining jar/dishes	Coplin Jar	VWR	100500-232
		Staining Dishes	VWR	25608-906
<input type="checkbox"/>	Section dryer oven	Epredia High Capacity Section Dryer (<i>Or equivalent. Thermal cycler may also be used for section drying.</i>)	Fisher Scientific	A84600051
<input type="checkbox"/>	Blank Slides	Shandon ColorFrost Plus Slides 25 x 75 x 1 mm (Optional)	Thermo Fisher Scientific	6776214
		Fisherbrand Premier Plain Glass Microscope Slides (Optional)	Thermo Fisher Scientific	12-544-4
Additional Materials				
<input type="checkbox"/>	Waterbath or			

For FFPE Tissue Sections: Deparaffinization & Decrosslinking

Thermomixer with 2 ml adapter

Thermal Cycler (*see Recommended Thermal Cyclers*)

Slide drying rack

Fume Hood

Vortex

This list may not include some standard laboratory equipment.

Tips & Best Practices

Icons



Tips & Best Practices section includes additional guidance



Signifies critical step requiring accurate execution



Troubleshooting section includes additional guidance

General Reagent Handling


- Fully thaw reagents at indicated temperatures. Thoroughly mix reagents before use.
- When pipette mixing reagents, unless otherwise specified, set pipette to 75% of total volume.
- Promptly move reagents back to the recommended storage.

Pipette Calibration

- Follow manufacturer's calibration and maintenance schedules.

Xenium Slide

- Xenium slides include an imageable area outlined by a white line measuring 12 mm x 24 mm, with an available sample positioning area measuring 235 mm² (10.45 mm x 22.45 mm). The available sample positioning area will be referred to as the Sample Area for the remainder of this document.
- The Sample Area is surrounded by fiducials. Tissue sections are placed within the Sample Area without obstructing the fiducials. The imageable area includes the area within the fiducial frame + Sample Area.
- The Sample Area can accommodate as many tissue sections as can fit within the space. Ensure tissue sections (including wax) DO NOT overlap.
- An etched label denoting the Slide ID, Part, and Version numbers is located at the bottom of the slide. Tissue sections should be placed on labeled-side of slide.



Xenium Slide Handling

- Always wear gloves when handling slides.
- The bottom of the slide is indicated by the etched label, which should be readable when in the proper position.
- The tissue sections should always be placed within the Sample Area on etched label side of the slide.
- Hold the slide on the label. DO NOT touch the tissue sections or near fiducials.
- Minimize exposure of the slides to sources of particles and fibers.
- Keep the slide cassette flat on the bench when adding reagents to the Sample Area.
- Ensure that no absorbent surface is in contact with the reagents on the slide during incubation.
- When pipetting reagent onto a slide, avoid generating bubbles. Avoid pipetting directly onto the tissue.



Slide Storage

- Always store unused slides at -20°C in their original packaging and keep sealed. Once opened, slides should remain at room temperature in a desiccator and be used within one week.
- Store slides containing FFPE tissues at room temperature in a desiccator for up to four weeks (not in a sealed container).

Xenium Cassette



- The Xenium Cassette is a single use item.
- The Xenium Cassette encases the slide and creates a leakproof well for adding reagents.
- Place the slide in the Xenium Cassette only when specified.
- Inner and outer tabs on the bottom half of the Xenium Cassette are used for holding the slide in the cassette. Applying excessive force to the cassette may cause the slide to break.
- The Xenium Cassette is assembled manually. See [Xenium Cassette Assembly & Removal](#) instructions for details.
- The Xenium Cassette includes an attached Xenium Gasket. The Xenium Gasket corresponds to the Sample Area on the slides.

- The etched slide label is visible in the label window when properly assembled.
- Ensure that the Xenium Cassette and gasket are free of debris prior to assembly. If placing the top half of the cassette on a surface, ensure the gasket faces up so it does not collect debris.
- Visually inspect the gasket to ensure it is seated properly. If the gasket appears warped, the Xenium Cassette is still safe to use as long as the cassette can fully close.


Xenium Cassette Assembly

Ensure slide is completely dry before assembling into cassette.



Exercise caution when handling slide edges to prevent injury.

Place top and bottom halves of cassette on bench




Place Xenium slide with tissue side facing upwards into bottom half of cassette; ensure label is toward bottom of cassette



Press slide down into grooves of the bottom half of the cassette until it sits firmly in place



Secure clips of top half with tabs of bottom half (on both sides)



Apply even pressure on top of cassette until all clips click shut. Verify that clips are completely secured over tabs




Slides in images are representative.




Once cassette is assembled, DO NOT remove slide until after Xenium Analyzer imaging and decoding for optional H&E staining step.

Xenium Cassette Removal


Pull inner clips from inner tabs to detach top and bottom halves of cassette



Open cassette by continuing to lift inner clips upward




Hold slide by the label and lift slide out from bottom half



Slides in images are representative.

Reagent Addition to Wells

- Place assembled cassette flat on a clean work surface.
- Dispense and remove reagents along the side of the well without touching the tissue sections and without introducing bubbles.
- Always cover the Sample Area completely when adding reagents to the well. A gentle tap may help spread the reagent more evenly.



Reagent Removal from Wells

- Place assembled cassette flat on a clean work surface.
- Slightly tilt the cassette while removing the reagent.
- Place the pipette tip on the bottom edge of the well.
- Remove reagents along the side of the well without touching the tissue sections.
- Remove all liquid from the well in each step.




Xenium Cassette Lid Application & Removal

Application

- Place the Xenium Cassette flat on a clean work surface.
- Hold the Xenium Cassette Lid with index and middle finger on two upper tabs and thumb on the lower clip.
- Align the Xenium Cassette Lid with the surface of the Xenium Cassette. Hook the two upper clips into the two holes on the top of the cassette.
- Push the lid down until the lower clip clicks into place.
- Inspect the lid to confirm placement.

Removal

- Place the Xenium Cassette flat on a clean work surface.
- Push on the top of the two upper tabs with index and middle fingers.
- Use thumb to push in on the lower clip.
- While maintaining inward pressure, pull upward with thumb until the lower clip disengages.
- Ensure that no liquid splashes out of the well.



Note that Xenium Cassette Lids are a single use item and should be discarded after each use. PBS-T washes DO NOT require sealing of the cassette with a lid.

Slide Incubation Guidance

Incubation at a specified temperature


Incubation using a Section Dryer Oven:

- Place the slides in a slide drying rack sideways to minimize paraffin wax entering neighboring tissue.
- Close the lid when incubating the slide in the oven.



Incubation using a Thermal Cycler:

- Position a Xenium Thermocycler Adaptor on a thermal cycler that is set at the incubation temperature. Ensure thermal cycler has reached appropriate temperature prior to starting incubation.
- Ensure that the Thermocycler Adaptor is fully inserted into the thermal cycler and is in contact with the thermal cycler block surface uniformly.
- When incubating a slide, position the slide on the Thermocycler Adaptor with the tissue side facing up.
- Ensure the Sample Area is aligned with the corresponding area on the Thermocycler Adaptor. DO NOT close lid.
- When incubating a slide encased in a Xenium Cassette, place the assembled unit on the Thermocycler Adaptor with the well facing up. The cassette should always be sealed with a Xenium Cassette Lid when on the Thermocycler Adaptor unless indicated otherwise.




Tightening the thermal cycler lid

- Thermal cycler lid contact with the Xenium Cassette Lid is critical for assay performance.
- Tighten the thermal cycler lid until an audible click is heard.
- Tightening past the click risks breaking the slide.


Incubation at room temperature

- Place the slide/Xenium Cassette on a flat, clean, non-absorbent work surface.
- Ensure that no absorbent surface is in contact with the reagents on the slide during incubation.

Processing a Single Xenium Slide

- Xenium reagent kits are sufficient for two reactions, and for optimal Xenium Analyzer throughput, two slides should be run at the same time.
-  It is possible to perform the Xenium In Situ Gene Expression workflow with a single slide. To do this, ensure the following best practices are followed for optimal assay performance:
 - Assemble a mock Xenium Cassette using a blank slide and a cassette from the Xenium Cassette Kit (2 cassettes), PN-1000566.
 - Insert the blank slide into the Xenium Cassette. Cassettes should be assembled following the instructions in [Tips & Best Practices for Xenium Cassette Assembly](#).
 - Attach a Xenium Cassette Lid from the Xenium Cassette Kit (2 cassettes), PN-1000566 to the cassette containing the blank slide following [Tips & Best Practices for Xenium Lid Application](#). It is not necessary to add liquid to the slide well before adding the lid.
 - For all incubation steps with the thermal cycler lid closed, ensure the mock slide cassette is placed alongside the Xenium slide cassette containing tissue on the Thermocycler Adaptor.

Tissue Detachment on Xenium Slides

-  Monitor section adhesion on Xenium slides throughout the workflow.
- Tissue detachment during the workflow can negatively impact performance. If observed, contact support@10xgenomics.com.
- For more information, refer to [Troubleshooting](#).

Protocol Steps & Timing

~4.5 hours

Steps	Timing	Stop & Store
Deparaffinization & Decrosslinking		
1.1	Buffer Preparation - Deparaffinization & Decrosslinking	30 min
1.2	Deparaffinization (includes 2 h baking at 60°C)	3 h
1.3	Cassette Assembly	10 min
1.4	Decrosslinking	1 h




Note there are no safe stopping points during this workflow.


1. Deparaffinization & Decrosslinking


1.0 Overview

This chapter provides guidance on deparaffinization and decrosslinking of Xenium slides containing FFPE tissue sections that are dried overnight in a desiccator. After paraffin is removed from the tissue sections during the deparaffinization process, tissues are rehydrated, and subsequently decrosslinked to release the sequestered RNA from the tissue. Step 1 (Probe Hybridization) of Xenium In Situ Gene Expression - Probe Hybridization, Ligation & Amplification User Guide (CG000582) should be immediately performed following deparaffinization and decrosslinking.



Protocol Overview





Get Started - Deparaffinization & Decrosslinking

Each 10x Genomics reagent tube is good for two Xenium Slides.

Deparaffinization Items	10x PN	Preparation & Handling	Storage	
Obtain				
<input type="checkbox"/>	Xylene	-	-	Ambient
<input type="checkbox"/>	Ethanol	-	Prepare Ethanol dilutions using Nuclease-free water.	Ambient
<input type="checkbox"/>	Nuclease-free Water	-	-	Ambient
<input type="checkbox"/>	10X PBS	-	-	Ambient
<input type="checkbox"/>	Forceps	-	-	Ambient
<input type="checkbox"/>	Slide Rack	-	-	Ambient
<input type="checkbox"/>	Coplin jars/Staining dishes	-	-	Ambient
<input type="checkbox"/>	Xenium Slides (2 pack) with FFPE tissue sections	3000941	Prepared according to Xenium In Situ for FFPE - Tissue Preparation Guide (CG000578).	Room temperature in a dessicator

Decrosslinking Items		10x PN	Preparation & Handling	Storage	
Equilibrate to room temperature					
<input type="checkbox"/>		Perm Enzyme B	3000533	Maintain at room temperature. Pipette mix, centrifuge briefly. DO NOT vortex.	-20°C
<input type="checkbox"/>		Xenium FFPE Tissue Enhancer	2000798	Thaw in a thermomixer for 30 min at 37°C, 300 rpm with shaking. Cool to room temperature for 5 min. Vortex for 30 sec and centrifuge briefly. Alternatively, thaw in a waterbath for 30 min at 37°C. Cool to room temperature for 5 min. Vortex for 30 sec and centrifuge briefly.*	-20°C
Obtain					
<input type="checkbox"/>		Nuclease-free Water	-	-	Ambient
<input type="checkbox"/>		1X PBS	-	Prepared at Step 1.1a.	Ambient
<input type="checkbox"/>		10% Tween-20	-	-	Ambient
<input type="checkbox"/>		Urea	-	-	Ambient



*Pre-heat thermomixer or waterbath to 37°C in advance of intended use.

1.1 Preparation - Buffers

For Deparaffinization:

Prepare all buffers fresh according to the tables below.

- a. Prepare 1X PBS. Add reagents in the order listed. Invert gently to mix. Maintain at room temperature. This volume of 1X PBS is sufficient for washes in all subsequent steps of this Demonstrated Protocol.

1X PBS				
Items		Stock	Final	Total Amount (ml)
<input type="checkbox"/>	Nuclease-free water	-	-	9.0
<input type="checkbox"/>	RNase free PBS	10X	1X	1.0
<input type="checkbox"/>	Total	-	-	10.0

- b. Prepare eight total coplin jars for deparaffinization steps. Prepare Ethanol dilutions using Nuclease-free water.

For Deparaffinization		
Items	Preparation & Handling	
<input type="checkbox"/>	Xylene	Label two coplin jars as Xylene Jar 1 and 2. Fill to capacity with xylene in each.
<input type="checkbox"/>	100% Ethanol	Label two coplin jars as 100% Ethanol Jar 1 and 2. Fill to capacity with 100% ethanol.
<input type="checkbox"/>	96% Ethanol	Label two coplin jars as 96% Ethanol Jar 1 and 2. Fill to capacity with 96% ethanol.
<input type="checkbox"/>	70% Ethanol	Label one coplin jar as 70% Ethanol Jar. Fill to capacity with 70% ethanol.
<input type="checkbox"/>	Nuclease-free water	Label one coplin jar as Nuclease-free Water Jar. Fill to capacity with Nuclease-free water.



Alternatively, a slide staining dish can be used in place of a coplin jar. Adjust volumes accordingly. Use xylene-resistant dishes, gloves, and forceps during workflow. Prepare fresh reagents after every 20 slides or every week (whichever comes first).

For Decrosslinking:

Prepare all buffers fresh according to the tables below.

- c. Using 1X PBS from step 1.1a, prepare PBS-Tween (PBS-T). Add reagents in the order listed. Invert gently to mix. Maintain at room temperature. This volume of PBS-T is sufficient for washes in all subsequent steps of this Demonstrated Protocol.

PBS-T				
Items		Stock	Final	Total Amount (µl)
<input type="checkbox"/>	1X PBS	-	-	4,975.0
<input type="checkbox"/>	Tween-20	10%	0.05%	25.0
<input type="checkbox"/>	Total	-	-	5,000.0

- d. Prepare Diluted Perm Enzyme B using 1X PBS prepared from step 1.1a. Add reagents in the order listed. Mix thoroughly with a 1-ml pipette set to 600 µl. Maintain at room temperature.

Diluted Perm Enzyme B				
Items		Stock	Final	Total Amount (µl)
<input type="checkbox"/>	1X PBS	1X	-	998.0
<input type="checkbox"/>	Perm Enzyme B <i>(Maintain at room temperature. Pipette mix, centrifuge briefly. DO NOT vortex).</i>	-	-	2.0
<input type="checkbox"/>	Total	-	-	1,000.0

- e. Prepare Decrosslinking Buffer using Diluted Perm Enzyme B prepared at step 1.1d. Add reagents in the order listed. Pipette mix thoroughly. Maintain at room temperature in the dark.

Decrosslinking Buffer					
Items		Stock	Final	1 slide+10% (µl)	2 slides+10% (µl)
<input type="checkbox"/>	Xenium FFPE Tissue Enhancer <i>(Thaw in a thermomixer for 30 min at 37°C, 300 rpm with shaking. Cool to room temperature for 5 min. Vortex for 30 sec and centrifuge briefly. Alternatively, thaw in a waterbath for 30 min at 37°C. Cool to room temperature for 5 min. Vortex for 30 sec and centrifuge briefly).</i>	-	-	508.8	1,017.5
<input type="checkbox"/>	Urea	8 M	0.5 M	34.4	68.8
<input type="checkbox"/>	Diluted Perm Enzyme B	-	-	6.9	13.8
<input type="checkbox"/>	Total	-	-	550.0	1,100.0


1.2 Deparaffinization

Deparaffinization steps should be performed in a fume hood due to the hazardous nature of xylene. Xylene jars should be covered at all times to prevent evaporation.


- a. Retrieve the slide with tissue sections from the desiccator after overnight drying.

Remove any marker annotations on slide using a lint-free laboratory wipe and 100% Ethanol.

- b. Place slide in a Section Dryer Oven and incubate uncovered at **60°C** for **2 h**. Keep the oven lid closed during incubation.



Alternatively, place a Thermocycler Adaptor on a thermal cycler set at **60°C**. Place slide on the Thermocycler Adaptor with the tissue side facing up and incubate at **60°C** for **2 h**. **DO NOT** close the thermal cycler lid.




- c. Remove from the oven or thermal cycler and allow the slide to cool down to **room temperature** for **7 min.**



Optional: photograph the slide against a black background during 7 min cool down at room temperature. This image can be used for comparison purposes to identify tissue detachment downstream in the workflow. Work quickly as this is not a safe stopping point. See [Troubleshooting](#) for more details.



- d. Gently immerse slide in the Xylene Jar 1. Secure the jar cap to prevent xylene loss.



Hold slide at label with forceps for xylene immersion steps. When immersing slides in xylene, ensure that the tissue sections are completely submerged.

- e. Incubate for **10 min.**
- f. Gently immerse slide in the Xylene Jar 2 and incubate for **10 min.**
- g. Gently immerse slide in the 100% Ethanol Jar 1 for **3 min.**

Hold slide at label with forceps for ethanol immersion steps. When immersing slides in ethanol, ensure that the tissue sections are completely submerged.

- h. Gently immerse slide in the 100% Ethanol Jar 2 for **3 min.**
- i. Gently immerse slide in the 96% Ethanol Jar 1 for **3 min.**
- j. Gently immerse slide in the 96% Ethanol Jar 2 for **3 min.**
- k. Gently immerse slide in the 70% Ethanol Jar for **3 min.**
- l. Gently immerse slide in the Nuclease-free water Jar for **20 sec.**

1.3 Cassette Assembly

- a. Remove any remaining Nuclease-free water from the slide using a lint-free laboratory wipe. Dry back of slide completely and front of slide outside of Sample Area without touching or damaging the tissue sections. Place the slide in the Xenium Cassette.



Refer to [Tips & Best Practices](#) for guidance on Xenium Cassette Assembly.

- b. Add **500 µl** 1X PBS to the well.

If processing two slides, leave second slide in water until the first cassette has been assembled and PBS has been added.

- c. Proceed **immediately** to Decrosslinking.



1.4 Decrosslinking

Reagent addition and removal should be done carefully. Remove reagents along the side of the well without touching the tissue sections and without introducing bubbles.

- a. Place a Xenium Thermocycler Adaptor in the thermal cycler.
- b. Prepare the thermal cycler with the following incubation protocol and start the program.

Lid Temperature	Reaction Volume	Run Time
80°C	100µL	-
Step	Temperature	Time
Hold	22°C	Hold
Decrosslinking	80°C	00:30:00
Re-equilibrate	22°C	00:10:00
Hold	22°C	Hold

- c. Remove 1X PBS from step 1.3b.
- d. Add **500 µl** Decrosslinking Buffer along the side of the well to uniformly cover the tissue sections, without introducing bubbles. Tap Xenium Cassette gently to ensure uniform coverage.
- e. Apply a new Xenium Cassette Lid on the Xenium Cassette and place the cassette on the Thermocycler Adaptor at **22°C**. Close the thermal cycler lid.
- f. Skip Hold step and initiate Decrosslinking.



Start thawing reagents for Probe Hybridization during Decrosslinking incubation as indicated in the Xenium In Situ Gene Expression - Probe Hybridization, Ligation & Amplification User Guide (CG000582).

- g. Remove Xenium Cassette from the thermal cycler and place on a flat, clean work surface.
- h. Remove the Xenium Cassette Lid and using a pipette, remove all Decrosslinking Buffer from the well. Discard old Cassette Lids.
- i. Add **500 µl** PBS-T to the well.
Note that PBS-T washes DO NOT require sealing of the cassette with a lid.
- j. Incubate for **1 min** at **room temperature**.
- k. Carefully remove all PBS-T from the well using a pipette tip.

l. Repeat steps **i-k** two more times.

m. Add **500 µl** PBS-T to the well.



Optional: photograph the slide against a black background. This image can be used for comparison purposes to identify tissue detachment downstream in the workflow. Work quickly as this is not a safe stopping point. See [Troubleshooting](#) for more details.



n. Proceed **immediately** to Xenium In Situ Gene Expression - Probe Hybridization, Ligation & Amplification User Guide (CG000582).

Troubleshooting

Tissue Detachment on Xenium Slides

Tissue sections may detach from Xenium slides during on-slide workflows. Tissue adhesion to the slide is impacted both by tissue processing and tissue quality.

Careful tissue preparation is critical for adhesion in on-slide workflows. Consult Xenium In Situ for FFPE - Tissue Preparation Guide (Document CG000579) for tissue QC and sectioning best practices. Below are some additional best practices for minimizing detachment during on-slide workflows:

- Do not pipette directly onto the tissue.
- Gently add and remove reagents from the well. Forceful addition or removal of reagents can agitate tissue and lead to detachment.
- Avoid touching in and around the Sample Area of the Xenium slide.
- Work quickly and carefully during reagent addition and removal.

In addition to following best practices, it is possible to monitor section adhesion on Xenium slides throughout the workflow. Taking a photograph of the slide at the beginning of the on-slide workflow and comparing with post-assay workflow images can help identify whether tissue shape has changed significantly, an indication of detachment. Steps when slide photos can be taken are noted in the protocol. These QC images can be compared with the DAPI overview scan as part of the Web Summary file to see whether tissue morphology has changed in the workflow.

If tissue detachment occurs, send pictures to support@10xgenomics.com for further assistance.

Cassette Assembly Failure

Incorrect assembly of the Xenium Cassette with a Xenium slide can negatively impact assay performance. Always dry the front and the back of the slide completely using a lint-free laboratory wipe while avoiding touching or damaging of the tissue sections. Inspect the slide carefully to ensure it is seated fully within the cassette before assembly.

Example scenarios that may indicate incorrect Xenium Cassette assembly are described below:

- If a gap appears between the two halves of the cassette after assembly.
- If the cassette does not click shut or appears domed after assembly.
- If the Xenium Thermocycler Adaptor is wet following removal from the thermal cycler, indicating reagent leakage from the cassette.


If the cassette is incorrectly assembled, disassemble and reassemble the cassette as instructed in [Tips & Best Practices](#).

Incorrect cassette assembly as indicated by a gap between the two halves of the cassette

Correct cassette assembly



Incorrect cassette assembly



Document Revision Summary

Document Number	CG000580
Title	Xenium In Situ for FFPE – Deparaffinization & Decrosslinking Demonstrated Protocol
Revision	Rev B
Revision Date	April 2023
Specific Changes	Updated Recommended Thermal Cyclers. Added guidance on usage of Xenium Cassette Lids to Tips & Best Practices. Updated guidance on handling and storage conditions of Perm Enzyme B.
General Changes	Updated for general minor consistency of language and terms throughout.

© 2023 10x Genomics, Inc. (10x Genomics). All rights reserved. Duplication and/or reproduction of all or any portion of this document without the express written consent of 10x Genomics, is strictly forbidden. Nothing contained herein shall constitute any warranty, express or implied, as to the performance of any products described herein. Any and all warranties applicable to any products are set forth in the applicable terms and conditions of sale accompanying the purchase of such product. 10x Genomics provides no warranty and hereby disclaims any and all warranties as to the use of any third-party products or protocols described herein. The use of products described herein is subject to certain restrictions as set forth in the applicable terms and conditions of sale accompanying the purchase of such product. A non-exhaustive list of 10x Genomics' marks, many of which are registered in the United States and other countries can be viewed at: www.10xgenomics.com/trademarks. 10x Genomics may refer to the products or services offered by other companies by their brand name or company name solely for clarity, and does not claim any rights in those third-party marks or names. 10x Genomics products may be covered by one or more of the patents as indicated at: www.10xgenomics.com/patents. The use of products described herein is subject to 10x Genomics Terms and Conditions of Sale, available at www.10xgenomics.com/legal-notices, or such other terms that have been agreed to in writing between 10x Genomics and user. All products and services described herein are intended FOR RESEARCH USE ONLY and NOT FOR USE IN DIAGNOSTIC PROCEDURES.

The use of 10x Genomics products in practicing the methods set forth herein has not been validated by 10x Genomics, and such non-validated use is NOT COVERED BY 10X GENOMICS STANDARD WARRANTY, AND 10X GENOMICS HEREBY DISCLAIMS ANY AND ALL WARRANTIES FOR SUCH USE. Nothing in this document should be construed as altering, waiving or amending in any manner 10x Genomics terms and conditions of sale for the Chromium Controller or the Chromium Single Cell Controller, consumables or software, including without limitation such terms and conditions relating to certain use restrictions, limited license, warranty and limitation of liability, and nothing in this document shall be deemed to be Documentation, as that term is set forth in such terms and conditions of sale. Nothing in this document shall be construed as any representation by 10x Genomics that it currently or will at any time in the future offer or in any way support any application set forth herein.

Contact:Email: support@10xgenomics.com

10x Genomics

6230 Stoneridge Mall Road

Pleasanton, CA